CONTROL
DRILLING SERVICE 1987 LTD.

Certificate of Calibration

Certificate:	64414	**Rig:**	
Date:	August 15 2019	**PO:**	
Customer:	K & S POWER TONGS	**Test Medium:**	Oil
Model:	6" Tong Torque Gauge	**Ambient Temp:**	21°C
Manufacturer:	CDS	**Relative Humidity:**	51% RH
Serial No:	# 142	**Customer #**	

Load/Pressure Test

Should Be Ft. Lbs.	Actual Ft. Lbs.
0	0
3 750	3 750
7 500	7 500
11 250	11 250
15 000	15 000

Technician: Dan M.

Comments: Calibrated f/ 0-15 000 Ft.Lbs. f/ 31" Handle w/ 4.08 Cell

Test Gauge: JW 121212-21 Cal. Date: May 7th 2019

The test gauge is certified to meet and/or exceed the requirements as outlined in the API specifications 6A, twentieth edition, October 2010

We hereby certify the we calibrated the above instrument, and that it is calibrated to ASME B40-1 Criteria

5921 - 87A STREET, EDMONTON, ALBERTA T6E 5W6
TELEPHONE (780) 465-6006 FAX (780) 469-8872