

Certificate of Calibration

Certificate: 64153 **Rig:** _____
Date: May 15th 2019 **PO:** _____
Customer: K&S POWER TONGS **Test Medium:** Oil
Model: 6" Tong Torque Gauge **Ambient Temp:** 21°C
Manufacturer: CDS **Relative Humidity:** 51% RH
Serial No: # 00374 **Customer #** _____

Load/Pressure Test

Should Be FT.LBS.	Actual FT.LBS.
0	0
3 750	3 750
7 500	7 500
11 250	11 250
15 000	15 000

Technician: Dan M. *[Signature]*
Comments: Calibrated f/ 0-15 000 Ft.Lbs. f/ 31" Handlef/ 4.08 Cell
Test Gauge: JW 121212-21 Cal. Date: May 7th 2019

The test gauge is certified to meet and/or exceed the requirements as outlined in the API specifications 6A, twentieth edition, October 2010

We hereby certify the we calibrated the above instrument, and that it is calibrated to ASME B40-1 Criteria