



## Certificate of Calibration

Certificate: 65290 Rig: \_\_\_\_\_  
Date: May 13th 2020 PO: \_\_\_\_\_  
Customer: K & S POWER TONGS Test Medium: Oil  
Model: Tong Torque Gauge Ambient Temp: 21°C  
Manufacturer: CDS Relative Humidity: 51% RH  
Serial No: # 1693 Customer # \_\_\_\_\_

### Load/Pressure Test

Should Be FT.LBS.	Actual FT.LBS.
0	0
3 750	3 750
7 500	7 500
11 250	11 250
15 000	15 000

Technician: Dan M.   
Comments: Calibrated f/ 0-15 000 Ft.Lbs. f/ 31" Handle f/ 4.08 Cell  
Test Gauge: JW 121212-21 Calibration Date: Feb. 14 2020

The test gauge is certified to meet and/or exceed the requirements as outlined in the API specifications 6A, twentieth edition, October 2010

We hereby certify the we calibrated the above instrument, and that it is calibrated to ASME B40-1 Criteria