



LOAD CELL CALIBRATION CERTIFICATE

CERTIFICATE #: 1404045

DATE: November 25, 2021
CUSTOMER: AllTorque Inventory
ADDRESS: Red Deer County, AB

WORK ORDER:
MODEL: TC10K
SERIAL #: 1404045

ALLTORQUE CERTIFIES THAT THE EQUIPMENT USED TO PERFORM THE TESTS ARE TRACEABLE TO PRIMARY STANDARDS AT NIST. THESE DOCUMENTS ARE AVAILABLE BY REQUEST

TEST 1

TEST 2

Table with 4 columns: APPLIED FORCE (LBS, FT/LBS), OUTPUT (LBS, FT/LBS) for TEST 1 and TEST 2. TEST 1 data includes values like 1045/1043, 2020/2021, etc.

TECHNICIAN: Maureen Tongol

OUTPUT ERROR (%): < +/- 0.5%

APPROVED: Wayne McLravey

TEST RESULT (PASS/FAIL): PASS

