

Certificate of Calibration

Certificate:	69420	Rig No.	· ·
Date:	25-Jul-23	Purchase Order:	
Customer:	K&S Power Tongs	Test Medium:	Oil
Model:	6" Tong Torque Gauge	Ambient Temp:	21°C
Manufacturer:	CDS	Relative Humidity:	51% RH
Serial No:	M14270	Customer #	

Load/Pressure Test

Should Be Ft. Lbs	Actual Ft. Lbs	
0	0	
7,500	7,500	
15,000	15,000	
22,500	22,500	
30,000	30,000	

Technician:

D.M

Comments

Calibrated for 26" handle and 4.4301 compression load cell

Test Gauge:

JW060516-21

Calibration Date:

May 6,2023

This gauge is certified to meet and/or exceed the requirements as outlined in API Specification 6A, Twenty first Edition in compliance with Section 10.2, Paragraph 10.2.1, 10.2.2 Subparagraph 10.2.2.1, 10.2.2.2, and 10.2.2.3, also spec 16A, forth Edition

April 2017, Section 7, Paragraph 7.1, 7.2 Subparagraph 7.2.1 and 7.2.2

We hereby certify that on the above date we repaired and recalibrated the above Instrument and that it calibrated to manufacturers specifications. The manufacturers specifications state an accuracy of +/- 2% of full scale calibration. The test gauge is calibrated in compliance with (CSA) CAN3-3-85 and A.N.S.I. B40-1. All test equipment is traceable to N.I.S.T. Standards, Customer Standards Document QCP 03-013 rev. 25

5921-87A Street, Edmonton, Alberta T6E 5W6 Telephone (780) 465-6006 Fax (780) 469-8872